

## TECHNICAL DATA

# Fluke IRR1-SOL Solar Irradiance Meter



### HIGH PRECISION MONO-CRYSTALLINE SOLAR SENSOR

Instantaneous irradiance measurements up to 1400 W/m<sup>2</sup>

### TWO OPTIONS FOR TEMPERATURE MEASUREMENT

Use the built-in temperature sensor or the external suction mount temperature probe to measure ambient and panel temperature

### INTEGRATED COMPASS

Measure and document roof or site orientation

### INCLINATION SENSOR

Know roof and PV panel tilt when surveying, installing, or adjusting the installation

**Make the critical measurements needed for installing, testing, maintaining, and reporting on solar panels or photovoltaic systems with one, easy-to-use tool.**

The Fluke IRR1-SOL Irradiance Meter has been designed from the ground up to simplify the installation, commissioning, and troubleshooting of photovoltaic arrays, measuring irradiance, temperature, inclination and direction of the solar array in a single handheld tool. With a rugged, compact design, a protective carrying case, and an easy-to-read, high-contrast LCD screen to read measurements in direct sunlight, the IRR1-SOL can go where you go. The simple user interface, instantaneous solar irradiation measurements and built-in temperature sensor make it easy to meet the IEC 62446-1 requirements for testing, documenting, and maintaining photovoltaic systems. Additionally, the integrated compass and inclination sensor allow you to quickly measure and document roof and site orientation, pitch, and panel tilt while surveying, installing, or adjusting an installation.

Whether working on a roof-mounted system or on a large field installation, the IRR1-SOL is the single-handed solution that every solar installer and technician needs in their tool bag.

### Use the IRR1-SOL for:

#### Photovoltaic system design and surveying

To find the expected production at a site, determine your solar resource while taking shading into account. The solar resource is measured in peak sun hours: the number of hours per day with 1,000 watts generated per square meter of solar array. Location, time of day, season, and weather conditions all influence peak sun hours. Use the Fluke IRR1-SOL to determine the actual solar irradiance (Watts/m<sup>2</sup>) and shading at the site to develop a baseline.

#### Measuring

Once your system is installed, make sure it is operating as designed by measuring its electrical characteristics and the actual power output of the array. The performance of a photovoltaic array is based on its current-voltage (IV) curve. Use the IRR1-SOL to obtain the amount of solar irradiance necessary to calculate the IV curve of the power output.

#### Comparing and diagnosing

Even when installed correctly, a photovoltaic system may not be producing the expected electrical output. In order to produce the expected output the system needs to receive the correct amount of irradiance energy to generate the DC voltage that is fed into the inverter.



## Specifications

Irradiance	
Measuring Range	0 to 1400 W/m <sup>2</sup>
Resolution	1 W/m <sup>2</sup>
Measuring Accuracy	± (5 % + 5 Digit)
Temperature Measurement	
Measuring range (°C)	-30 °C to 100 °C
Resolution	0.1 °C
Measuring Accuracy	±1 °C @ -10 °C to 75 °C, ±2 °C @ -30 °C to -10 °C and 175 °C to 100 °C

Note: Temperature measurement response time: ~30 sec.

Inclination Angle	
Measuring Range	-90° to +90°
Resolution	0.1°
Measuring Accuracy	± 1.5° @ -50° to +50°, ±2.5° @ -85° to -50° and +50° to +85° ±3.5° @ -90° to -85° and +85° to +90°

Compass	
Measuring Range	0° to 360°
Resolution	1°
Measuring Accuracy	± 7°

Note: a) Measurements valid for device inclination between -20° and +20° to horizontal. Outside that range on LCD will be shown "----".  
b) Result is referred to magnetic north.

Temperature	
Operating Temperature IRR1-SOL	-20 °C to 50 °C (humidity <80 %), noncondensing
Operating Temperature 80PR-IRR	-30 °C to 100 °C
Storage Temperature	-30 °C to 60 °C (humidity <80 %)
Altitude	0 m to max. 2000 m

Electromagnetic Compatibility (EMC)	
International	<p>IEC 61326-1: Portable Electromagnetic Environment CISPR 11: Group 1, Class A Group 1: Equipment has intentionally generated and/or uses conductively-coupled radio frequency energy that is necessary for radio frequency energy that is necessary for the internal function of the equipment itself. Class A: Equipment is suitable for use in all establishments other than domestic and those directly connected to a low voltage power supply network that supplies buildings used for domestic purposes. There may be potential difficulties in ensuring electromagnetic compatibility in other environments due to conducted and radiated disturbances. Caution: This equipment is not intended for use in residential environments and may not provide adequate protection to radio reception in such environments.</p>

## Specifications continued

Korea (KCC)	Class A Equipment (Industrial Broadcasting & Communication Equipment) Class A: Equipment meets requirements for industrial electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and not to be used in homes.
USA (FCC)	47 CFR 15 subpart B. This product is considered an exempt device per clause 15.103.
<b>Protection</b>	
IP Protection	IP40
<b>Power Supply &amp; Battery Life</b>	
Batteries	4 AA Alkaline Batteries
Battery Life (typical)	50 hours (> 9000 readings)
Auto Power Off	30 minutes
<b>Dimensions</b>	
L x W x H	150 x 80 x 35 mm
Weight	231 g

## Ordering information

### Fluke IRR1-SOL Solar Irradiance Meter

Includes: FLK-IRR1-SOL Solar Irradiance Meter, FLK-80PR-IRR External Temperature Probe with Suction Cup, C250 Carrying Case with Shoulder Strap, (4) AA Alkaline Batteries, User Manual.



**Fluke.** *Keeping your world up and running.®*

**Fluke Corporation**  
PO Box 9090, Everett, WA 98206 U.S.A.

**Fluke Europe B.V.**  
PO Box 1186, 5602 BD  
Eindhoven, The Netherlands

**For more information call:**  
**Fluke Australia**  
Unit 16/7 Anella Avenue  
Castle Hill, NSW, 2154  
Australia  
Phone: 1300 1 FLUKE (35853)  
Fax: +61 2 8850 3300  
Email: [auinfo@fluke.com](mailto:auinfo@fluke.com)  
Website: [www.fluke.com.au](http://www.fluke.com.au)

©2020 Fluke Corporation.  
Specifications subject to change without notice.  
10/2020 200400-en

**Modification of this document is not permitted without written permission from Fluke Corporation.**