

EQUIVALENT SOUNDABSORPTION AREA PER OBJECT conform EN-ISO 354:2003



principal: BuzziSpace

#16; BuzziFalls Large 250x120 cm, hanging



Absorb, versie 5.10.1 mode 9, PM:JK, file: a4435 E#:1-36 F#:594-629 A#:630 T<sub>1</sub>=21,2 °C T<sub>2</sub>=21,1 °C p<sub>1</sub>=101,6 kPa p<sub>2</sub>=101,6 kPa h<sub>1</sub>=66,7 % h<sub>2</sub>=65,0 %

volume reverberation room: 214 m<sup>3</sup>


number of objects measured: 2

measured at: Peutz Laboratory for Acoustics

signal: broad-band noise

bandwidth: 1/3 octave

- 1/3 oct.
- \* 1/1 oct.



	0,02	0,41	0,97	1,85	3,01	3,71
1/3 oct.	0,16	0,60	1,27	2,26	3,30	3,71
	0,28	0,85	1,65	2,46	3,58	4,00
1/1 oct.	<b>0,15</b>	<b>0,62</b>	<b>1,30</b>	<b>2,19</b>	<b>3,30</b>	<b>3,81</b>

RA

publication is permitted for the entire page only

Mook, measured at 17-08-2023